

**GE Healthcare** 

## Technical Publications

Document Number: DOC1731688 Revision: 2

# DoseWatch

Version 2.0

## Integrating the Healthcare Enterprise (IHE) Integration Statement

Copyright© 2015 by General Electric Company

### INTRODUCTION

#### OVERVIEW

This IHE Integration Statement describes the intended conformance of this GE product with the IHE Technical Framework. The IHE Technical Framework defines a subset of the functional components of the healthcare enterprise, called IHE Actors, and specifies their interactions in terms of a set of coordinated, standards-based transactions. These transactions are organized into functional units called Integration Profiles that highlight their capacity to address specific clinical needs.

Users familiar with these concepts can use Integration Statements as an aid to determine what level of integration this product supports with other systems and what clinical and operational benefits such integration might provide.

Other GE Integration Statements as well as further information on IHE may be found on Internet at: <u>http://www.ge.com/ihe</u>

#### IMPORTANT REMARKS

The use of these IHE Integration Statements, **by itself, it is not sufficient to ensure that inter-operation will be successful**. The **user** needs to proceed with caution and ensure that a qualified systems' integration (such as that provided by GE's IT Professional Services) address at least three issues:

- Integration The integration of any system into an integrated department or enterprise requires an analysis of the applications requirements. The design of a solution that integrates GE systems with non-GE systems is the user's responsibility and should not be underestimated. Special care should be exercised in defining the expected workflow, the special conditions, and the clinical practice to be supported to ensure a safe and effective operation.
- Validation Testing the complete range of possible interactions between any GE system and non-GE system, before the integration is declared operational, should not be overlooked. Therefore, the **user** should ensure that any non-GE provider accepts full responsibility for all validation required for their connection with GE systems. This includes the accuracy of the data once it has crossed the interface and the stability of the data for the intended applications. Such a validation is required before any clinical use (diagnosis and/or treatment) is performed.
- Future Evolution GE understands that the IHE Technical Framework and the underlying communication standards will evolve to meet the user's growing requirements. GE is actively involved in the development of the IHE Technical Framework and the underlying standards. Such evolutions may require changes to already installed systems. In addition, GE reserves the right to discontinue or make changes to product features including those reflected by this IHE Integration Statement. The user should ensure that any non-GE system that connects with GE systems also plans for the future evolution of IHE. Failure to do so will likely result in the loss of function and/or connectivity as the IHE Integration profiles evolve and GE Products are enhanced to support these changes.

### **IHE INTEGRATION STATEMENT**

General Electric - IHE Integration Statement			ent	Date	
				5 Apr 2014	
Vendor		Product Name	Version		
GE Healthcare		DoseWatch	V2.0		
This product impleme Integration Profiles, A		L ctions required in the IHE Technical Framewo ions listed below:	ork to support t	he IHE	
Integration Profiles Implemented		Actors Implemented	Options Implemented		
Access to Radiology Information		Image Display	None	None	
Consistent Time		Time Client	None	None	
Patient Administration Management		Patient Demographics Consumer	Merge	Merge	
Patient Identifier Cross- referencing for MPI		Patient Identifier Cross-reference Consumer	PIX Update	PIX Update Notification	
Radiation Exposure Monitoring		Dose Information Consumer	None		
		Dose Information Reporter	None		
		Dose Register	None		
Scheduled Workflow.b		Image Display	None		
		Performed Procedure Step Manager	None		
		re IHE information: http://www.ge.com/ih			
L	inks to Stand	ards Conformance Statements for the Impl	ementation		
DICOM	http://www3.gehealthcare.com/~/media/documents/us- global/products/interoperability/dicom/dose-management/gehc-dicom- conformance_dosewatch-2-0_doc1742008_rev1.pdf				
		Links to general information on IHE			
		IHE International: www.ihe.net			

This page is left blank

#### Manufacturer:

GE Medical Systems - SCS (A General Electric Company, going to market as GE Healthcare) 8, place de l'Hôpital 67000 Strasbourg FRANCE

+33 3 88 22 80 40



www.gehealthcare.com