

GE Healthcare

Technical Publications

Direction: 5342851

Revision: 01

LOGIQ P5

Integrating the Healthcare Enterprise (IHE) Integration Statement

Copyright© 2009 by General Electric Company

INTRODUCTION

OVERVIEW

This IHE Integration Statement describes the intended conformance of this GE product with the IHE Technical Framework. The IHE Technical Framework defines a subset of the functional components of the healthcare enterprise, called IHE Actors, and specifies their interactions in terms of a set of coordinated, standards-based transactions. These transactions are organized into functional units called Integration Profiles that highlight their capacity to address specific clinical needs.

Users familiar with these concepts can use Integration Statements as an aid to determine what level of integration this product supports with other systems and what clinical and operational benefits such integration might provide.

Other GE Integration Statements as well as further information on IHE may be found on Internet at: **<u>ge.com/ihe</u>**

IMPORTANT REMARKS

The use of these IHE Integration Statements, by itself, it is not sufficient to ensure that inter-operation will be successful. The user needs to proceed with caution and ensure that a qualified systems' integration (such as that provided by GE's IT Professional Services) address at least four issues:

- Integration The integration of any system into an integrated department or enterprise requires an analysis of the applications requirements. The design of a solution that integrates GE systems with non–GE systems is the user's responsibility and should not be underestimated. Special care should be exercised in defining the expected workflow, the special conditions, and the clinical practice to be supported to ensure a safe and effective operation.
- Validation Testing the complete range of possible interactions between any GE system and non–GE system, before the integration is declared operational, should not be overlooked. Therefore, the user should ensure that any non–GE provider accepts full responsibility for all validation required for their connection with GE systems. This includes the accuracy of the data once it has crossed the interface and the stability of the data for the intended applications. Such a validation is required before any clinical use (diagnosis and/or treatment) is performed.
- Future Evolution GE understands that the IHE Technical Framework and the underlying communication standards will evolve to meet the user's growing requirements. GE is actively involved in the development of the IHE Technical Framework and the underlying standards. Such evolutions may require changes to already installed systems. In addition, GE reserves the right to discontinue or make changes to product features including those reflected by this IHE Integration Statement. The user should ensure that any non–GE system that connects with GE systems also plans for the future evolution of IHE. Failure to do so will likely result in the loss of function and/or connectivity as the IHE Integration profiles evolve and GE Products are enhanced to support these changes.

IHE INTEGRATION STATEMENT

General Electric - IHE Integration Statement			tement	Date	
				Oct 30,2008	
Vendor		Product Name		Version	
GE Healthcare		LOGIQ P5		R3.0.X	
	mplements all tra	nsactions required in the IHE Technical Fra ted below:	amework to support	t the IHE Integration	
Integration Profiles Implemented		Actors Implemented	Opti	Options Implemented	
Scheduled Workflow		Acquisition Modality	Broad Wor	Broad Worklist Query	
			Patient Bas	Patient Based Worklist Query	
		Image Display	None	None	
PIR		Acquisition Modality	None	None	
ED		Evidence Creator	None		
ARI		Image Display None			
Internet add		althcare IHE information: http://www.g		· ·	
HL7	Not Applicabl	ble			
DICOM	http://www.ge 100_Rev3.0	gehealthcare.com/usen/interoperability/dicom/docs/LOGIQ_P5_version_3.x.x_517804 .0.pdf			
		Links to general information or	n IHE		
In North America: www.ihe.net		In Europe: <u>www.ihe-europe.org</u>	In Japan: <u>www.jira-net.or.jp/ihe-j</u>		
www.him	ss.org/ihe				

This page is left blank



GE Healthcare

GE Healthcare – Americas : Fax 262-544-3384 P.O.Box 414, Milwaukee, Wisconsin, 53201 U.S.A Internet – http://www.gehealthcare.com

GE Healthcare – Europe : Fax 33-1-30-70-95-95 Paris, France Internet - http://www.gehealthcareeurope.com

GE Healthcare – Asia Tokyo, Japan : Fax 81-3-3223-8560 Singapore : Fax 65-291-7006