GE Healthcare

Innova 2100-IQ, 3100/3100-IQ, 4100/4100-IQ Integrating the Healthcare Enterprise (IHE) Integration Statement



Revision History

Date	Part/Rev	Change Description
Nov. 20, 2008	5330553-1-8EN rev 1	Initial release.
Mar 18, 2010	5330553-2-8EN rev 1	In table, the line "DICOM" has been modified.

Introduction

Overview

This IHE Integration Statement describes the intended conformance of this GE product with the IHE Technical Framework. The IHE Technical Framework defines a subset of the functional components of the healthcare enterprise, called IHE Actors, and specifies their interactions in terms of a set of coordinated, standards-based transactions. These transactions are organized into functional units called Integration Profiles that highlight their capacity to address specific clinical needs.

Users familiar with these concepts can use Integration Statements as an aid to determine what level of integration this product supports with other systems and what clinical and operational benefits such integration might provide.

Other GE Integration Statements as well as further information on IHE may be found on internet at: **ge.com/ihe**.

Important remarks

The use of these IHE Integration Statements, by itself, it is not sufficient to ensure that interoperation will be successful. The user needs to proceed with caution and ensure that a qualified systems ' integration (such as GE's IT Professional Services) address at least four issues:

- Integration The integration of any system into an integrated department or enterprise
 requires an analysis of the applications requirements. The design of a solution that
 integrates GE systems with non–GE systems is the user's responsibility and should not be
 underestimated. Special care should be exercised in defining the expected workflow, the
 special conditions, and the clinical practice to be supported to ensure a safe and effective
 operation.
- Validation Testing the complete range of possible interactions between any GE system
 and non-GE system, before the integration is declared operational, should not be
 overlooked. Therefore, the user should ensure that any non-GE provider accepts full
 responsibility for all validation required for their connection with GE systems. This includes
 the accuracy of the data once it has crossed the interface and the stability of the data for the
 intended applications. Such a validation is required before any clinical use (diagnosis and/or
 treatment) is performed.
- Future Evolution GE understands that the IHE Technical Framework and the underlying communication standards will evolve to meet the user's growing requirements. GE is actively involved in the development of the IHE Technical Framework and the underlying standards. Such evolutions may require changes to already installed systems. In addition, GE reserves the right to discontinue or make changes to the support of communications features (on its products) reflected on by these IHE Integration Statements. The user should ensure that any non–GE system that connects with GE systems also plans for the future evolution of IHE. Failure to do so will likely result in the loss of function and/or connectivity as the IHE Integration profiles evolve and GE Products are enhanced to support these changes.

www.jira-net.or.jp/ihe-j

www.himss.org/ihe

Conoral Floatric IUE Integration Statement		
ne integration Statement	Nov 20, 2008	
Product Name	Version	
INNOVA	2100-IQ, 3100/3100-IQ, 4100/4100-IQ	
nsactions required in the IHE Technical	Framework to support the IHE	
d below:		
Actors Implemented	Options Implemented	
Acquisition Modality	Broad Worklist	
	PPS Exception Management	
	Patient Based Query	
Acquisition Modality	Broad Worklist	
	PPS Exception Management	
	Patient Based Query	
Acquisition Modality	None	
Time Client	None	
I Systems IHE information: ge.com/ihe	2	
ce Statements for the Implementation		
HL7 Not Applicable		
To be released		
) IHE		
	INNOVA INNOVA INNOVA Insactions required in the IHE Technical Insactions below: Actors Implemented Acquisition Modality Acquisition Modality Acquisition Modality Time Client I Systems IHE information: ge.com/ihe Ce Statements for the Implementation Not Applicable	

GE Medical Systems, a General Electric Company, going to market as GE Healthcare 283, rue de la Minière 78530 Buc FRANCE

www.gehealthcare.com

