

Technical Publications

Document Number: DOC0687367

Revision: 1

Vivid E9
Version 10

Integrating the Healthcare Enterprise (IHE) Integration Statement

INTRODUCTION

OVERVIEW

This IHE Integration Statement describes the intended conformance of this GE product with the IHE Technical Framework. The IHE Technical Framework defines a subset of the functional components of the healthcare enterprise, called IHE Actors, and specifies their interactions in terms of a set of coordinated, standards-based transactions. These transactions are organized into functional units called Integration Profiles that highlight their capacity to address specific clinical needs.

Users familiar with these concepts can use Integration Statements as an aid to determine what level of integration this product supports with other systems and what clinical and operational benefits such integration might provide.

Other GE Integration Statements as well as further information on IHE may be found on Internet at: **ge.com/ihe**

IMPORTANT REMARKS

The use of these IHE Integration Statements, by itself, it is not sufficient to ensure that inter-operation will be successful. The user needs to proceed with caution and ensure that a qualified systems' integration (such as that provided by GE's IT Professional Services) address at least four issues:

- Integration The integration of any system into an integrated department or enterprise requires an analysis of the applications requirements. The design of a solution that integrates GE systems with non—GE systems is the user's responsibility and should not be underestimated. Special care should be exercised in defining the expected workflow, the special conditions, and the clinical practice to be supported to ensure a safe and effective operation.
- Validation Testing the complete range of possible interactions between any GE system and non-GE system, before the integration is declared operational, should not be overlooked. Therefore, the user should ensure that any non-GE provider accepts full responsibility for all validation required for their connection with GE systems. This includes the accuracy of the data once it has crossed the interface and the stability of the data for the intended applications. Such a validation is required before any clinical use (diagnosis and/or treatment) is performed.
- Future Evolution GE understands that the IHE Technical Framework and the underlying communication standards will evolve to meet the user's growing requirements. GE is actively involved in the development of the IHE Technical Framework and the underlying standards. Such evolutions may require changes to already installed systems. In addition, GE reserves the right to discontinue or make changes to product features including those reflected by this IHE Integration Statement. The user should ensure that any non–GE system that connects with GE systems also plans for the future evolution of IHE. Failure to do so will likely result in the loss of function and/or connectivity as the IHE Integration profiles evolve and GE Products are enhanced to support these changes.

IHE INTEGRATION STATEMENT

Genera	tement	Date				
0.011010		16 Feb 2010				
Vendor		Product Name	V	Version		
GE Healthcare		Vivid E9		10		
This product implem Profiles, Actors and		ns required in the IHE Technical Frame ow:	work to support the II	HE Integration		
Integration Profiles Implemented		Actors Implemented	Options	Options Implemented		
Echo Workflow		Acquisition Modality	Stress	Stress		
Evidence Documents		Acquisition Modality	Echo Evide	Echo Evidence		
		re IHE information: ge.com/ihe				
Lii	nks to Standard	ds Conformance Statements for th	e Implementation			
HL7	N/A					
DICOM	www.gehealthcare.com/usen/interoperability/dicom/docs/DOC0687366_R1.pdf					
	L	inks to general information on IH	E			
In North America: www.ihe.net		In Europe: www.ihe-europe.net	In Japan: www.jira-net.or.jp/ihe-j			

This page is left blank



GE Healthcare

GE Healthcare – Americas : Fax 262-544-3384 P.O.Box 414, Milwaukee, Wisconsin, 53201 U.S.A Internet – http://www.gehealthcare.com

GE Healthcare – Europe : Fax 33-1-30-70-95-95 Paris, France Internet - http://www.gehealthcareeurope.com

GE Healthcare – Asia Tokyo, Japan : Fax 81-3-3223-8560 Singapore : Fax 65-291-7006